

赤外レーザヘテロダイン分光器 MILAHI を用いた超高分解能惑星大気観測 Ultra-high resolution observations of planetary atmospheres using Mid-Infrared LAsEr Heterodyne Instrument (MILAHI)

中川 広務^{1*}, 青木 翔平¹, 笠羽 康正¹, 村田 功¹, 黒田 壮大¹, 岡野 章一¹, Guido Sonnabend²
NAKAGAWA, Hiromu^{1*}, AOKI, Shohei¹, KASABA, Yasumasa¹, MURATA, Isao¹, Morihiro Kuroda¹, OKANO, Shoichi¹,
Guido Sonnabend²

¹ 東北大学・理, ² ケルン大学

¹Geophys., Tohoku University, ²University of Cologne

We have developed a ultra-high spectral resolution spectrometer, called Mid-Infrared LAsEr Heterodyne Instrument (MILAHI). It is for the applications to astronomy and planetary atmospheric science in 7-11 μm wavelength at a spectral resolution resolution of up to 10^{7-8} and a bandwidth of 1GHz. We just finish the development phase of this project.

High-resolution spectroscopy in the mid-infrared regime is a versatile tool for studies of planetary atmospheres. With the highest possible spectral resolution provided by heterodyne techniques, fully resolved molecular features enables us to retrieve many physical parameters from single lines migrated in strong absorption of terrestrial atmospheric molecule bands. Because many key species in the planetary atmosphere are also abundant in the terrestrial atmosphere, high-resolution directly leads to less ambiguity. It also allows us to measure slow wind velocities with the order of 10-100m/s directly.

The heterodyne spectroscopy has been developed by our group from 1980s, in order to detect minor constituents in the terrestrial atmosphere [Taguchi et al., 1990]. The renovation with a wide-band detector, the quantum-cascade (QC) lasers and CO₂ gas laser allows us to apply this instrument to tiny planetary atmosphere.

Our performance achieved the proper level for this target. (1) System noise: At 10.3 μm , we achieved 3000 K (NEP of 2.24 W/Hz^{1/2} at 3MHz resolution). It leads to a minimal detectable brightness temperature difference of 37mK within 10min at 1.5 MHz bandwidth, corresponds to a minimum flux difference of 0.48 ergs/(scm²cm-1Sr) for extended source. (2) Spectral resolution: It can be achieved to be 20 MHz with a feedback using gas-cell absorption spectra.

The telluric CO₂ and O₃ absorption spectra had been obtained from the sunlight background in the lab at Sendai. On January in 2012, our equipment was mounted on the Higashi-Hiroshima 1.5m telescope, and succeeded to detect the telluric O₃ spectra obtained from moonlight. We also aimed Venus and standard stars. Unfortunately, the final success was prevented by bad weathers, but the S/N gained by these target told us that we should get the Venus and Mars spectrum with this design.

Now, we try to refine the emission spectra of the QC lasers, which provide us very wide tuneability (5cm⁻¹, and 20cm⁻¹) to operate the heterodyne system.

Although a telescope dedicated to this instrument does not exist yet, we expect to attach it to the PLANETS telescope at the top of Mt. Haleakala at Hawaii, which is now in development by PPARC / Tohoku Univ. with IfA / Univ. Hawaii (USA), Kiepenheuer Inst. f. Sonnen. (Germany), Univ. Nac. Aut. de Mexico, Univ. Turku (Finland), Harlinton Inovative Optics Co. (USA), Stan Truitt Breckenridge Astronomical Ltd (USA), and collaborators. Its first light is, if all things are going well, in 2014.

キーワード: 分光, 赤外, 惑星大気, 超高分解能, ヘテロダイン, レーザ

Keywords: spectroscopy, infrared, planetary atmospheres, high spectral resolution, heterodyne, laser