

宇宙機での同位体計測を目指したマルチターン・飛行時間計測型質量分析器 A multi-turn time-of-flight isotope analyzer for space application

横田 勝一郎^{1*}, 豊田岐聡², 青木順², 栗原 純一³, 斎藤 義文¹

Shoichiro Yokota^{1*}, Michisato Toyoda², Jun Aoki², Junichi Kurihara³, Yoshifumi Saito¹

¹ 宇宙科学研究所, ² 大阪大学, ³ 北海道大学

¹ISAS/JAXA, ²Osaka Univ., ³Hokkaido Univ.

In order to study terrestrial or planetary plasma environment in situ low-energy ion measurements are indispensable and thus have been done by a variety of ion analyzers. Detailed studies of plasma characteristics demand mass analyses as well as energy analyses. In case of measuring a variety of ions originating from planetary atmospheres, we need to measure the ion composition with high mass resolution. Although we have achieved the measurements of the ion composition by mass analyzers around planetary environment, higher mass resolution is now needed in order to distinguish heavy species and isotopes. For the future isotope measurements around moons, planets and asteroids, we are developing a high-mass-resolution mass analyzer. One of our scientific objects is to measure the Martian atmospheric escape and evolution. Mass resolution (m/dm) of 100 is generally needed for the isotope analysis of planetary particles. However the Martian atmospheric escape and evolution science requires $m/dm > 3,000$ to discriminate N_2 from CO .

Low-energy particle measurement group of ISAS has developed a time-of-flight (TOF) ion mass analyzer with mass resolution of about 20 for KAGUYA, which succeeded in measuring ions originating from the lunar exosphere and surface. It is also preparing a TOF mass analyzer with mass resolution of 40 for the BepiColombo mission. Multi-turn TOF mass spectrometers (MULTUM), where ions are stored in a fixed orbit within electrostatic sectors and allowed to propagate the same orbit numerous times, have been developed by Osaka Univ. mass spectrometry group. One of the MULTUM series achieves the mass resolution over 30000 with the size of 20cm x 20cm. We have prepared a test model of the ion optics of the isotope analyzer which employs the MULTUM technique. We are also developing a pulsive high voltage power supply (HVPS) for the test model of the ion optics. We will report test results of the MULTUM optics and the HVPS performance.

キーワード: 質量分析, 同位体, 惑星探査

Keywords: mass analysis, isotope, planetary exploration