

LA-MC-ICPMS による Sr 同位体測定の新しいバイアス補正法 A new analytical bias correction for in-situ Sr isotope analysis of plagioclase crystals using LA-MC-ICPMS

木村 純一^{1*}, 高橋 俊郎¹, 常 青¹
Jun-Ichi Kimura^{1*}, Toshiro Takahashi¹, Qing Chang¹

¹ 海洋研究開発機構

¹IFREE/JAMSTEC

A new analytical protocol was developed for correcting baseline-induced biases during the analysis of Sr isotope ratios in plagioclase using excimer laser-ablation Aridus-dual-intake-system multiple-collector inductively coupled plasma mass spectrometry (LA-MC-ICP-MS). Residual analytical biases of $^{84}\text{Sr}/^{86}\text{Sr}$ and $^{87}\text{Sr}/^{86}\text{Sr}$ were observed after applying on-peak background subtractions and mass-fractionation corrections using the conventional $^{86}\text{Sr}/^{88}\text{Sr}$ exponential internal normalization. The residual biases occurred only for samples analysed with LA and not for solution analyses using Aridus with the same instrumental setup. Based on observations from the ablation of NaCl and olivine crystals, we concluded that this was due to suppression and enhancement of the Kr baseline by loading of the LA sample aerosols and by the introduction of Kr from the samples, respectively. We also found that both the $^{84}\text{Sr}/^{86}\text{Sr}$ and $^{87}\text{Sr}/^{86}\text{Sr}$ isotope ratios were affected proportionally by the baseline biases of the LA analyses of an isotopically/compositionally homogeneous anorthite plagioclase (MkAn) from Miyakejima, Japan, and similar results were seen in theoretical calculations. Therefore, the bias correction for the target $^{87}\text{Sr}/^{86}\text{Sr}$ ratios was available using the shift in the simultaneously measured $^{84}\text{Sr}/^{86}\text{Sr}$ from the natural ratio. We then determined the correlation factors between $^{84}\text{Sr}/^{86}\text{Sr}$ and $^{87}\text{Sr}/^{86}\text{Sr}$ by analyzing MkAn, which reproduced the theoretical factors obtained from numerical simulations. By applying the new correction protocol, the accurate measurement of unknown plagioclase samples was possible. This was confirmed by comparisons with the results of micro-milling thermal ionization mass spectrometry (MM-TIMS) of the same plagioclase crystals with various compositions. The new correction technique improved reproducibility by a factor five, providing a basis for an accurate Sr isotope analysis using LA-MC-ICP-MS.

キーワード: Sr 同位体, レーザーアブレーション, MC-ICPMS, バイアス補正
Keywords: Sr isotope, laser ablation, MC-ICPMS, bias correction